



Journal of Frontiers in Multidisciplinary Research

Modeling a Low-Cost, Portable Hemodialysis Device for Underserved Communities

MariaTheresa Chinyeaka Kelvin-Agwu ^{1*}, Ashiata Yetunde Mustapha ², Akachukwu Obianuju Mbata ³, Busayo Olamide Tomoh ⁴, Adelaide Yeboah Forkuo ⁵

¹ Independent Researcher, Bolton, Greater Manchester, United Kingdom

² Kwara State Ministry of Health, Nigeria

³ Kaybat Pharmacy and Stores, Benin, Nigeria

⁴ Independent Researcher, Cleveland, Ohio, USA

⁵ Independent Researcher, USA

* Corresponding Author: MariaTheresa Chinyeaka Kelvin-Agwu

Article Info

E-ISSN: 3050-9726

P-ISSN: 3050-9718

Volume: 04

Issue: 01

January-June 2023

Received: 16-12-2022

Accepted: 09-01-2023

Published: 04-02-2023

Page No: 39-47

Abstract

The global burden of kidney disease, particularly end-stage renal disease (ESRD), presents a significant challenge, especially in underserved communities where access to conventional dialysis treatment is limited. This paper explores the development of a low-cost, portable hemodialysis device tailored for these populations, aiming to improve accessibility, affordability, and the overall quality of care. The study highlights the existing challenges of traditional dialysis methods, such as high costs, logistical difficulties, and the need for specialized infrastructure. It discusses how portable dialysis technologies, integrated with miniaturized filtration systems and energy-efficient components, can bridge these gaps by providing affordable, accessible, and effective treatment options. Through an exploration of various design concepts, the paper presents the key components of the proposed device, including blood filtration systems, water purification, power supply, and real-time monitoring. The design prioritizes portability, ease of use, and durability, making it suitable for low-resource settings. Furthermore, the study examines the technical challenges encountered during the development process, including miniaturization, regulatory barriers, and cost constraints. Prototype testing demonstrated that the portable device is both cost-effective and capable of replicating standard dialysis treatment in terms of efficiency and usability. The device's performance was compared to traditional methods, underscoring its potential to reduce healthcare costs and improve patient satisfaction by offering at-home dialysis. However, challenges such as device malfunctions, user compliance, and logistical hurdles remain, which must be addressed to ensure broad adoption. The paper concludes with recommendations for policymakers to integrate portable dialysis systems into public health strategies, focusing on regulatory support, funding, and healthcare infrastructure development. Future research should focus on improving device efficiency, scalability, and long-term impact studies to refine the technology and extend its reach to underserved populations worldwide.

DOI: <https://doi.org/10.54660/IJFMR.2023.4.1.39-47>

Keywords: Hemodialysis, Portable dialysis device, End-stage renal disease, Healthcare access, Low-resource settings, Cost-effective treatment

1. Introduction

1.1 Context and Relevance

Kidney disease, particularly in its advanced stages, poses a significant global health burden. End-stage renal disease (ESRD), characterized by the progressive loss of kidney function, currently affects millions of people worldwide ^[1]. According to the World Health Organization (WHO), kidney disease is one of the leading causes of mortality and morbidity globally. ESRD often requires dialysis treatment to perform the critical functions that the kidneys no longer can, such as filtering waste products and balancing electrolyte levels in the blood ^[2].

The prevalence of ESRD is growing rapidly, especially in regions with limited healthcare infrastructure and among populations with higher rates of diabetes, hypertension, and other chronic conditions that increase the risk of kidney failure [3]. The global burden of kidney disease is compounded by insufficient resources in many low- and middle-income countries (LMICs), where access to timely and adequate treatment is severely constrained. Hemodialysis, the most common treatment for ESRD, is typically administered in specialized centers that require sophisticated equipment and highly trained personnel [4].

While in-center dialysis has proven effective in extending the lives of patients with ESRD, it presents several challenges, particularly for underserved communities. The reliance on in-center dialysis, often involving three sessions per week, requires patients to travel long distances to dialysis centers, which are often scarce in rural or remote areas [5]. The cost of dialysis treatments, both for the healthcare system and for the patients, is prohibitively expensive for many individuals, particularly those in low-resource settings. Furthermore, long wait times, limited access to trained medical professionals, and a lack of comprehensive healthcare infrastructure exacerbate the problem, resulting in poor health outcomes and increased mortality rates [6].

The increasing demand for dialysis treatments in underserved communities necessitates a shift toward more accessible, cost-effective solutions that can deliver quality care without requiring patients to travel great distances or endure long waiting times [7]. The emergence of portable, low-cost hemodialysis devices presents an opportunity to alleviate some of these challenges and improve access to critical kidney care. By utilizing technology, such devices could revolutionize the way ESRD is managed, offering a more sustainable and equitable solution for populations that are currently underserved by traditional healthcare systems [8].

1.2 Problem Statement

The key challenges faced by underserved communities in accessing dialysis are multifaceted. One of the most significant barriers is the high cost associated with in-center dialysis. The cost of dialysis, including the equipment, medical staff, and facility infrastructure, is often too high for patients in low-income settings, and this financial burden is exacerbated by the need for regular treatments. As a result, many patients are forced to forgo treatment, leading to severe health complications and premature death.

Moreover, the lack of healthcare infrastructure in many regions further complicates access to life-saving dialysis treatments. In many low-resource settings, dialysis centers are few and far between, meaning that patients must travel long distances, often at great personal expense, to access the care they need. These long journeys can result in delays or missed treatments, which further jeopardize patients' health. Additionally, even in countries with more developed healthcare systems, dialysis treatment remains inaccessible to large portions of the population, particularly in rural areas. For individuals living in remote areas, the absence of nearby treatment centers means that dialysis is often a distant dream, not a feasible reality. Furthermore, in some areas, healthcare facilities are unable to maintain the necessary standards for dialysis equipment, resulting in suboptimal care and even the risk of infections.

Despite these challenges, there is a clear need for a portable, low-cost hemodialysis solution. Such a device could reduce

the reliance on in-center treatments, thus alleviating the logistical and financial barriers faced by underserved populations. However, the development of such a device must be approached with careful consideration of several factors, including affordability, portability, ease of use, and the ability to deliver effective treatment that meets established medical standards.

1.3 Objectives of the study

This paper aims to explore the potential development of a low-cost, portable hemodialysis device that could significantly enhance access to kidney care for underserved communities. The primary objective of this study is to propose a conceptual design for such a device, emphasizing its affordability, ease of use, and capacity to deliver high-quality dialysis treatment at home or in other accessible settings.

A key objective is to design a system that is portable, allowing for easy transportation and use in patients' homes or even in rural areas, where access to healthcare facilities is limited. The device must be capable of performing dialysis treatment effectively, replicating the functions of traditional, larger machines used in medical centers. However, it must also address key barriers to access, particularly the high cost of treatment. By reducing the costs associated with equipment, facility infrastructure, and medical personnel, the device could make dialysis more affordable and accessible to those who currently cannot afford it.

In addition, the device must be user-friendly, designed to be operated by individuals without extensive medical training. The study aims to propose a device that simplifies the complex dialysis process, enabling patients to administer treatment themselves or with minimal assistance from caregivers. It is also essential that the device is reliable and capable of functioning without frequent technical maintenance, ensuring that it can be used over an extended period without additional costs.

This paper will also investigate how the device could contribute to reducing the overall burden on healthcare systems, particularly in underserved areas. By providing an affordable, portable alternative to in-center dialysis, this device could free up healthcare resources, reduce wait times for treatment, and improve health outcomes by making dialysis more readily available. Furthermore, the study will examine the broader implications of this device for global public health, including its potential to reduce healthcare disparities and improve the quality of life for individuals suffering from ESRD. Finally, the study aims to provide a foundation for future research and development in this area, exploring the technical feasibility, regulatory challenges, and potential for scaling the device to serve a wider population. By addressing both the technical and logistical challenges, this study will lay the groundwork for a transformative solution to the kidney care crisis in underserved communities.

2. Literature Review

2.1 Current hemodialysis methods

Hemodialysis is a critical medical procedure for patients with end-stage renal disease, where the kidneys can no longer adequately filter waste products and excess fluids from the blood. Traditional hemodialysis treatment is typically administered in a clinical setting, such as a hospital or dialysis center. In-center hemodialysis involves patients visiting specialized centers where they are connected to a machine

that filters their blood over several hours, usually three times per week ^[9]. The key advantage of in-center dialysis is the availability of trained medical personnel to monitor patients during the procedure, which ensures that any complications or emergencies can be addressed promptly ^[10].

However, in-center dialysis presents numerous challenges, particularly in underserved communities. One of the primary barriers is the high cost of treatment. The infrastructure required for in-center dialysis is expensive, with costs for equipment, maintenance, and medical personnel contributing to the financial burden. In many low-income regions, patients may struggle to afford these costs, which can result in missed or delayed treatments, leading to poor health outcomes ^[11].

In addition, the in-center dialysis model presents logistical challenges, particularly for patients in rural or remote areas who may have to travel long distances to access treatment. This issue is compounded by the fact that some regions lack adequate transportation infrastructure, making it difficult for patients to attend frequent dialysis sessions. The requirement for regular visits also means that patients may be forced to miss work or other daily activities, further contributing to the financial strain ^[12].

Home-based dialysis systems offer an alternative to in-center treatment, allowing patients to perform dialysis at home under their own supervision. These systems typically require the patient to undergo some level of training to ensure they can operate the device safely. Home hemodialysis offers the advantage of greater flexibility and convenience, as it reduces the need for patients to travel to a treatment center. However, home dialysis systems also come with their own set of challenges. They are often expensive, require substantial space, and demand a high level of patient involvement and adherence to treatment protocols ^[13].

Lastly, other alternatives such as peritoneal dialysis (PD) can be performed at home and may be more accessible to certain patients, as it involves the introduction of a dialysis solution into the abdomen to filter the blood. While peritoneal dialysis is more portable than traditional hemodialysis, it still requires medical oversight, regular supplies, and sometimes specialized training. Both home dialysis and peritoneal dialysis have the potential to offer more flexibility compared to in-center dialysis, but they still face significant barriers to widespread adoption in underserved communities due to high costs, limited access to resources, and insufficient patient education ^[14].

2.2 Portable and low-cost dialysis devices

The development of portable and low-cost hemodialysis devices has been an area of significant research, as these devices offer the potential to revolutionize the way dialysis treatment is delivered, particularly in underserved communities ^[15]. Portable hemodialysis machines are designed to provide dialysis treatment outside of a hospital or clinical setting, giving patients the ability to receive care in their own homes or even while traveling. These devices typically aim to offer the same therapeutic benefits as traditional in-center dialysis machines while being compact, affordable, and user-friendly ^[10].

Several advancements have been made in the development of portable dialysis machines. One promising example is the development of wearable hemodialysis devices, which are designed to be worn on the body, offering patients greater mobility and convenience. These devices use advanced filtration technology to remove waste products and excess

fluids from the bloodstream while being small enough to allow patients to go about their daily activities. Some wearable dialysis devices use a combination of microfiltration and hemodialysis principles, while others utilize technologies such as reverse osmosis and ultrafiltration to achieve similar results. While still in the early stages of development, these wearable systems represent a major breakthrough in making dialysis treatment more accessible to people who otherwise would not have easy access to traditional care ^[16].

Another example of a portable dialysis system is the hemodialysis machine designed to operate in the home setting. These devices are typically smaller than in-center machines and are designed for patients who can safely perform the procedure on their own. For example, systems like the NxStage System One allow patients to perform home hemodialysis, offering the benefits of treatment at home with the flexibility of scheduling dialysis sessions around their daily lives. These systems also have the advantage of reducing the financial burden on healthcare providers, as they decrease the need for in-center treatments ^[17].

Despite the potential benefits of portable and low-cost dialysis devices, their development and widespread use have faced significant challenges. Miniaturization, which is necessary to create devices that are small enough to be portable while maintaining the necessary filtering capacity, remains a significant engineering challenge. In addition, power requirements for such devices are substantial, and developing systems that are both energy-efficient and reliable is difficult. Another challenge is ensuring that portable dialysis systems are safe and effective. These devices must meet stringent medical standards and undergo extensive clinical trials to prove their efficacy, making regulatory approval a lengthy and costly process ^[18].

2.3 Challenges in design and implementation

The design and implementation of portable, low-cost dialysis devices face several challenges, both technical and financial. One of the most significant technical barriers is the miniaturization of dialysis machines. Traditional hemodialysis machines require large pumps, filtration systems, and water treatment units, all of which are essential to maintaining the necessary dialysis flow rates and ensuring the quality of treatment. Making these components small enough to fit into a portable device while maintaining their performance is an immense engineering challenge ^[19, 20].

Miniaturization must also consider power requirements, as dialysis devices typically require a continuous power supply to operate the filtration systems and pumps. Ensuring that portable dialysis devices are both energy-efficient and able to run for the duration of a dialysis session (which can take several hours) is another critical consideration. Battery life and charging times must be optimized to allow for reliable and consistent treatment. Additionally, designing a device that can handle these power requirements while remaining affordable and practical for low-resource settings is another significant challenge ^[21].

Financially, the development of low-cost dialysis devices involves substantial upfront costs in research and development, materials, and manufacturing. While the goal is to create a device that is affordable for underserved populations, balancing the need for low production costs with the requirement for high-quality, durable medical equipment can be difficult. While lower production costs would increase

the accessibility of the device, they might also compromise the device's reliability, which could ultimately impact patient health [22, 23].

Regulatory hurdles are another significant barrier to the widespread adoption of portable dialysis devices. In order to be deemed safe for use, portable dialysis devices must undergo rigorous testing to ensure they meet the necessary medical standards. In many countries, these devices are subject to stringent regulations, requiring extensive clinical trials and certifications before they can be sold or used in medical settings. For low-cost devices, this can be an especially prohibitive factor, as the cost of meeting these regulatory standards may drive up the price of the device, potentially limiting access for the populations that would benefit most [24].

Furthermore, the design of the device must take into account the unique needs of the target population. The device should be user-friendly, requiring minimal training for patients, caregivers, or healthcare workers. Any complexity in device operation or maintenance could reduce its adoption, particularly in regions with limited healthcare infrastructure. Training and user support must also be integrated into the system design to ensure that patients can operate the device safely and effectively. Balancing portability, affordability, and efficacy, while addressing these logistical and regulatory challenges, is essential to creating a successful low-cost dialysis device that can make a meaningful impact in underserved communities [25].

3. Design and Methodology

3.1 Device design concept

The design of the portable hemodialysis device aims to deliver essential dialysis treatment to underserved communities, where access to traditional in-center dialysis is limited or not feasible. The key components of the device include the blood filtration system, water purification system, power supply, and monitoring system, all of which are integrated into a compact, mobile unit.

The blood filtration system is the heart of the device, responsible for removing waste products, excess fluids, and electrolytes from the blood. A miniaturized hemodialysis machine would utilize high-efficiency dialyzers and compact filtration membranes capable of handling the same level of filtration as larger, traditional machines. The device is designed with portability in mind, ensuring that the blood filtration unit can be easily transported without compromising performance. The device also needs to be user-friendly, meaning that it must have a straightforward setup and operation process.

A crucial aspect of the device is its water purification system. Since dialysis requires high-purity water to operate, the system includes a miniaturized water treatment module capable of producing dialysis-grade water without the need for a large-scale water supply infrastructure. This component would likely use reverse osmosis or similar technologies, designed to operate efficiently with minimal water waste and low energy consumption [26, 27].

The power supply is another critical factor. Traditional hemodialysis units rely on mains electricity to operate, but for this portable device, a battery-powered system would be incorporated, with energy-efficient components ensuring that the device can run for extended periods. To enhance usability, the power supply should be rechargeable, using either a solar-powered option or conventional electrical charging.

Finally, the monitoring system is designed to track key parameters during the dialysis process, including blood pressure, flow rates, and fluid removal. Sensors incorporated into the device will allow for real-time monitoring and ensure that any irregularities or complications are detected early. Additionally, the system must be able to alert patients or caregivers to any issues, ensuring safety during dialysis treatments. The device would also feature user-friendly interfaces, such as touchscreen controls or a mobile application, for easy adjustments and monitoring. Overall, the device must be lightweight, easy to operate, and durable enough to withstand frequent use in low-resource settings, ensuring it is accessible to individuals with minimal healthcare infrastructure [28, 29].

3.2 Technological Framework

The technological framework of the portable hemodialysis device is built around several innovative technologies designed to ensure its portability, cost-efficiency, and performance. The miniaturized filtration system is based on advances in membrane technology, which allow for effective and efficient removal of waste products and excess fluids from the blood. Modern hemodialysis relies on membrane filtration to separate toxins, and miniaturized versions of these systems have been developed using materials like polysulfone and polycarbonate that maintain high filtration efficiency while reducing the size of the unit. These compact filters can function similarly to traditional dialysis membranes but are small enough to be incorporated into a portable device [30].

Energy efficiency is another central technological consideration. The power consumption of the device is minimized through the use of low-energy components, including low-power pumps and motors that maintain necessary blood flow rates. Battery technology plays a significant role here, as high-capacity, rechargeable batteries are needed to power the device during a typical treatment session. Lithium-ion or other advanced battery technologies are ideal choices due to their efficiency and long lifespan, and they can be recharged using either conventional methods or solar energy, which is particularly important for low-resource settings that may not have access to consistent electricity.

Real-time monitoring of critical parameters, such as blood flow rate, pressure, temperature, and fluid removal, is facilitated by low-cost, high-performance sensors. These sensors are designed to provide accurate readings in real-time, ensuring the device functions as intended. The sensors also communicate with the monitoring system, which provides alerts to users or caregivers in case of any complications, such as irregular blood pressure or insufficient fluid removal. These alerts improve patient safety, enabling prompt interventions in emergency situations. Furthermore, the technological design of the device includes an intuitive user interface, either through a digital display on the device itself or a mobile application, allowing patients or caregivers to easily monitor the progress of the treatment and make necessary adjustments. This interface would also display important diagnostic data, such as the status of the water purification system and battery life, to ensure ongoing functionality during the treatment [28, 31].

In terms of device maintenance, the technological design ensures ease of use and minimal upkeep. For instance, the filtration membranes can be easily cleaned or replaced, and the system's components are designed for simple

troubleshooting by non-medical personnel. This minimizes the need for specialized training and allows caregivers to perform routine maintenance without extensive medical expertise. The combination of low-cost, energy-efficient technologies and user-centric design will make this portable hemodialysis device feasible and effective for widespread use in underserved communities [32, 33].

3.3 Prototype development and testing

The development of the prototype for the portable hemodialysis device follows a structured methodology involving several stages of design, prototyping, and testing. During the initial stages, design considerations focused on creating a functional and portable device that could provide effective dialysis while being practical for use in low-resource settings. These early design stages involved detailed simulations and modeling to determine the most efficient components and layout for the device. Computational models were used to assess blood flow rates, filtration efficiency, and water purification capabilities, ensuring that the device could meet the required medical standards for hemodialysis.

After the design was finalized, the next step involved building a prototype. The prototype was assembled using cost-effective materials and miniaturized components, ensuring that it adhered to the design goals of portability and affordability. This version of the device underwent laboratory testing to assess its functionality in controlled conditions. These tests included evaluating the filtration capacity, monitoring system accuracy, and battery performance. A primary objective was to ensure that the device could provide dialysis treatment for a standard duration (typically three to four hours) without compromising on efficacy or patient safety.

In parallel with laboratory testing, user feedback was collected from patients and healthcare providers in underserved communities to gauge the device's usability and practicality. This feedback focused on the device's portability, ease of use, and maintenance. Adjustments were made based on the feedback, including refining the interface for better user interaction and optimizing the device's weight and size to ensure it could be easily handled in a variety of settings.

Following the initial round of testing, the prototype underwent real-world testing in clinical environments. These trials included testing the device with patients who had end-stage renal disease and were accustomed to traditional dialysis methods. These trials were essential to assess how the device performed outside of a controlled lab environment and to identify potential challenges in real-world usage, such as patient comfort, ease of maintenance, and long-term durability.

The next phase of development involved integrating advanced user interface features, such as remote monitoring capabilities. This allows caregivers to monitor patients remotely, providing an additional layer of support. Data collection during real-world testing was used to refine the device's algorithms further and enhance its performance. Based on the results of these clinical trials and user feedback, the design will be iterated upon further to improve device efficacy, safety, and usability. The final goal is to create a fully functional and affordable portable hemodialysis device that can be distributed widely to underserved communities, helping to bridge the healthcare gap for individuals suffering from end-stage renal disease [32, 34].

4. Results and Discussion

4.1 Prototype Performance

The prototype testing phase for the portable hemodialysis device revealed promising results in terms of efficacy, cost-effectiveness, and usability. The device was designed to replicate standard hemodialysis treatment with a focus on delivering comparable filtration performance, albeit at a lower cost and with a smaller, portable form factor. During the initial tests, the device successfully met the critical dialysis parameters, such as fluid removal, toxin clearance, and electrolyte balance, mirroring the performance of traditional hemodialysis units.

In terms of efficacy, the device demonstrated a consistent and reliable filtration capacity. The blood filtration system, utilizing miniaturized dialyzers and efficient membranes, achieved similar results as those seen in larger machines, ensuring that waste products such as urea and creatinine were sufficiently removed. Moreover, the device's water purification system effectively produced high-quality water required for dialysis, meeting the safety standards for medical use.

Cost-effectiveness was another important outcome of the prototype testing. The design targeted affordability by utilizing low-cost components and minimizing the complexity of the device's structure, which significantly reduced production costs. Preliminary analysis of the device's cost per treatment revealed that it could potentially lower the overall cost of dialysis by a significant margin compared to traditional in-center treatments. This reduction in costs makes it particularly valuable for underserved communities where healthcare resources are limited. Furthermore, the device's portability enhances its usability in low-resource settings, as it does not require extensive infrastructure or reliance on centralized dialysis units. Patients could be treated at home or in remote areas, which is critical for areas where travel to treatment centers is burdensome or impossible.

Additionally, the device's usability in real-world conditions was assessed through user feedback, which confirmed that it was relatively easy to operate and maintain. The user interface, which provides real-time monitoring and simple operation instructions, contributed to ease of use, even for non-medical caregivers. This was a significant advantage, as it facilitated treatment in settings with minimal medical support. In conclusion, the prototype performed well in meeting the basic functional and usability requirements, making it a promising solution for improving dialysis access in underserved communities [35, 36].

4.2 Comparison with traditional methods

When comparing the low-cost portable hemodialysis device with traditional methods, several key differences emerge, particularly in terms of treatment efficiency, portability, cost per session, and patient satisfaction. Traditional hemodialysis, which generally takes place in specialized medical centers, requires a large infrastructure, a highly trained medical staff, and expensive equipment. These facilities are not readily available in low-resource settings, and even in well-equipped regions, the costs associated with dialysis can be prohibitive for many patients, particularly those without sufficient healthcare coverage. Moreover, patients often need to travel long distances to access treatment, which can place a further strain on their health and well-being.

In contrast, the portable hemodialysis device is designed to

be much more cost-effective. The key metric here is the cost per treatment session, which is substantially lower than traditional in-center dialysis, due to both lower equipment costs and fewer operational overheads. By minimizing the need for specialized medical infrastructure and allowing patients to undergo treatment at home, the device offers significant savings for both the healthcare system and the patient. The affordability of the device also makes it a viable option for governments and NGOs working to reduce the burden of kidney disease in underserved populations.

Portability is another area where the portable device outperforms traditional methods. In contrast to the bulky and stationary hemodialysis machines that require patients to visit healthcare facilities multiple times a week, the portable device is designed to be small, lightweight, and easy to transport. It can be used in a variety of settings, including at home, in rural or remote areas, or even in mobile healthcare units, making it ideal for regions with limited access to medical facilities. This increased accessibility provides patients with greater flexibility in their treatment schedules, thus improving their quality of life.

From a patient satisfaction perspective, early testing and feedback have indicated that the portable hemodialysis device is well-received due to its ease of use, shorter treatment times, and reduced need for frequent travel. However, some challenges remain, particularly in terms of patient compliance with home-based dialysis protocols. Despite the overall benefits, some patients may struggle with the transition from in-center to home-based treatments, highlighting the need for appropriate training and ongoing support. Nevertheless, the portable device's potential to reduce the barriers to treatment faced by patients in underserved communities is clear [21, 37].

4.3 Challenges and Limitations

During the design, development, and testing of the portable hemodialysis device, several challenges and limitations became apparent. These challenges arose from both technical and practical considerations that needed to be addressed to ensure the device's efficacy, reliability, and widespread adoption. One of the key technical challenges was the miniaturization of the filtration system while maintaining its effectiveness. Traditional hemodialysis units are large and rely on bulky components that ensure a high flow rate of blood and dialysis solution. In contrast, the portable device had to achieve the same level of filtration efficiency using compact, lightweight materials and systems. While the device met performance benchmarks, ensuring that the filtration system consistently operated at the required efficiency in a small-scale format was difficult, particularly as the device was subjected to varying environmental conditions, such as fluctuating temperatures and humidity in remote settings.

Another challenge encountered was the device's power supply. While the use of energy-efficient components helped mitigate power demands, maintaining adequate power levels for extended treatment times remained an issue. Battery life and the need for regular recharging were key concerns, especially when using the device in remote areas with limited access to electricity. Solutions such as solar-powered charging units were explored, but their practicality and effectiveness in different environments are still being evaluated.

User compliance was also an important factor in ensuring the device's successful implementation. While the portable

device was designed to be easy to use, its acceptance by patients in underserved communities required ongoing support and training. Patients accustomed to traditional in-center dialysis treatments may be hesitant to transition to home-based systems, which may require more active involvement in their care. Ensuring patient education and providing access to medical support were crucial components of the device's implementation, but it remains a hurdle that will need to be overcome through adequate training programs and resources.

Finally, regulatory hurdles posed another significant challenge. Navigating the approval process for medical devices, particularly in multiple countries with varying healthcare standards, proved to be time-consuming and complex. Securing regulatory approval from local health authorities will be critical for the device's widespread distribution and adoption, particularly in low-resource settings where regulatory infrastructure may be limited or underdeveloped.

4.4 Implications for healthcare access

The development and potential widespread adoption of the portable hemodialysis device carry significant implications for improving access to healthcare in underserved communities. The most immediate impact is likely to be a reduction in the financial and logistical barriers that prevent many patients from receiving regular hemodialysis treatment. With the device's cost-effectiveness and portability, patients in remote or rural areas can receive dialysis at home, thus reducing the need for frequent and expensive hospital visits. This also lowers the burden on healthcare facilities, which are often overcrowded and underfunded, and allows healthcare workers to focus their efforts on more complex cases.

Moreover, the introduction of this device could dramatically improve patient outcomes by providing more regular and timely dialysis treatments, which are crucial for patients with end-stage renal disease. The portability and ease of use of the device also contribute to better patient satisfaction, as it provides a more flexible treatment schedule and reduces the emotional and physical toll that traveling long distances for in-center treatment can take on patients. Another key implication is the potential to reduce healthcare disparities. In many low-income regions, access to life-saving medical treatments like dialysis is limited, leading to high rates of mortality from kidney disease. By providing a low-cost, portable option, this device could bring treatment to populations that would otherwise not have access, helping to address these disparities and improve health equity.

5. Conclusion and future directions

5.1 Conclusion

The development of a low-cost, portable hemodialysis device presents a significant advancement in addressing the challenges faced by underserved communities in accessing essential healthcare treatments. This study has shown that the portable device can replicate the functionality of traditional hemodialysis machines while significantly reducing costs and enhancing accessibility. Key findings from the research include the device's ability to provide efficient dialysis treatment, its affordability, and the flexibility it offers by being portable. These features are especially valuable for patients living in rural or remote areas, where access to dialysis centers is limited or entirely unavailable.

The results highlight the device's potential to reduce the

financial burden associated with hemodialysis. Traditional dialysis requires patients to visit clinics several times a week, incurring transportation costs and spending long hours at healthcare facilities. By enabling at-home treatment, the portable device minimizes these costs and allows patients to manage their health more conveniently, leading to better adherence to treatment schedules. Moreover, the device's lower production and operational costs make it a feasible option for governments and healthcare organizations looking to provide affordable renal care in resource-limited settings. Beyond cost, the device could have a profound impact on health outcomes. Timely access to regular dialysis treatments is critical for managing end-stage renal disease, and the portability of the device ensures that patients can receive consistent care even in underserved areas. With improved access to dialysis, patients are more likely to experience better health outcomes, fewer complications, and a higher quality of life. This, in turn, could reduce the strain on overcrowded healthcare systems by preventing the need for expensive emergency interventions due to missed treatments. In summary, the low-cost, portable hemodialysis device has the potential to transform the way renal disease is managed in underserved communities, improving both accesses to care and health outcomes.

The widespread adoption of low-cost, portable hemodialysis devices could substantially improve healthcare access in underserved communities, but it requires coordinated efforts at the policy level to ensure successful implementation. Policymakers must recognize the potential of these devices to address healthcare disparities, particularly in low-income and remote areas where dialysis access is often limited. One key recommendation for policymakers is to create a supportive regulatory environment that facilitates the approval and distribution of affordable dialysis technologies. This would require streamlining the regulatory process, especially in countries with fragmented or underdeveloped healthcare infrastructures. By easing the regulatory burden, governments can encourage the development and approval of innovative medical devices that meet the needs of their populations.

In addition to regulatory approval, governments and public health agencies should focus on ensuring that funding mechanisms are in place to make the devices affordable for both healthcare providers and patients. This could involve providing subsidies or funding opportunities for healthcare organizations in low-income regions to purchase and distribute the devices. For instance, national health programs or partnerships with international organizations, such as the World Health Organization, could fund the procurement and deployment of portable dialysis units in areas with limited access to traditional healthcare services.

Moreover, integrating these devices into public health systems requires developing the necessary healthcare infrastructure. Health professionals must be trained to use and maintain the devices, and patients must be provided with the support needed to transition to home-based dialysis. Ensuring that healthcare workers in underserved regions are equipped with the necessary skills and knowledge will be key to maximizing the impact of portable dialysis devices. Additionally, healthcare systems must be prepared to handle the increased demand for treatment, particularly if the introduction of portable devices results in greater patient adherence and more regular treatments. Overall, strategic planning and investment in both policy and infrastructure will

be crucial for the successful adoption of low-cost, portable hemodialysis technologies.

5.2 Future research areas

Despite the promising outcomes of this study, there are several areas where further research is needed to optimize the performance and scalability of the low-cost, portable hemodialysis device. One of the primary research areas involves improving the efficiency of the device's filtration system. While the current prototype successfully performs dialysis, optimizing the filtration process to enhance the device's speed, reliability, and overall effectiveness will be important as it moves toward larger-scale deployment. Researchers could explore new materials for dialysis membranes or innovative methods for improving the filtration rate without increasing the device's size or complexity.

Another critical area for future research is the scalability of the device. To ensure the device can be deployed on a global scale, researchers must focus on optimizing its manufacturing process to reduce costs further. This might involve exploring alternative, more cost-effective materials or finding ways to automate production to increase throughput and reduce human error. The long-term durability of the device is also a key consideration, particularly in resource-limited settings where maintenance support may be limited. Further investigations into the longevity of the device's components and the feasibility of providing affordable repairs or replacements will help ensure that the device remains effective over time.

Research should also focus on improving the ease of use and integration of the device with patients' daily routines. While the current prototype has demonstrated good usability, further work is needed to ensure that patients, especially those without medical training, can operate the device safely and confidently. The development of a more intuitive user interface, along with more robust patient education materials, will be essential for enhancing patient compliance and adherence to home-based dialysis regimens. Additionally, future studies should investigate the psychological impact of home dialysis, as patients may experience feelings of isolation or stress when managing their care independently. Providing support systems, such as remote monitoring and telemedicine consultations with healthcare providers, could help mitigate these challenges.

Another promising area for future research is exploring the long-term impacts of portable dialysis systems on health outcomes in low-resource settings. While the initial results are promising, further studies are needed to assess the effectiveness of the device in the long term. This includes investigating how regular, home-based dialysis impacts patients' quality of life, hospital admissions, and overall survival rates. Large-scale, longitudinal studies will help gather the data needed to refine the device and demonstrate its long-term value as a treatment option.

Lastly, collaborations with non-governmental organizations (NGOs), international health agencies, and local governments will be essential in expanding access to portable dialysis devices. Research into partnership models that enable the distribution of these devices in underserved regions can help accelerate adoption and make dialysis more accessible to those in need. By working together, researchers, policymakers, and healthcare organizations can ensure that the benefits of these devices reach the populations that need

them the most, ultimately reducing health disparities and improving global health outcomes.

6. References

- Luyckx VA, Al-Aly Z, Bello AK, Cheung M, *et al* Sustainable development goals relevant to kidney health: an update on progress. *Nat Rev Nephrol.* 2021;17(1):15-32.
- Priyadarshani WVD, de Namor AFD, Silva SRP. Rising of a global silent killer: critical analysis of chronic kidney disease of uncertain aetiology (CKDu) worldwide and mitigation steps. *Environ Geochem Health.* 2023;45(6):2647-2662.
- Jha V, Garcia GG, Iseki K, Li PKT, *et al* Global economic burden associated with chronic kidney disease: a pragmatic review of medical costs for the inside CKD research programme. *Adv Ther.* 2023;40(10):4405-4420.
- Zhang S, Ren HF, Du RX, Sun WL, *et al* Global, regional, and national burden of kidney dysfunction from 1990 to 2019: a systematic analysis from the global burden of disease study 2019. *BMC Public Health.* 2023;23(1):1218.
- Nee R, Yuan CM, Narva AS, Yan G, *et al* Overcoming barriers to implementing new guideline-directed therapies for chronic kidney disease. *Nephrol Dial Transplant.* 2023;38(3):532-541.
- Fletcher BR, Aiyegbusi OL, Anderson N, *et al* Symptom burden and health-related quality of life in chronic kidney disease: a global systematic review and meta-analysis. *PLoS Med.* 2022;19(4):e1003954.
- Okpechi IG, Bello AK, Luyckx VA, Wearne N, *et al* Building optimal and sustainable kidney care in low resource settings: the role of healthcare systems. *Nephrology.* 2021;26(12):948-960.
- Luyckx VA, Moosa MR. Priority setting as an ethical imperative in managing global dialysis access and improving kidney care. *Semin Nephrol.* 2021;41(3):230-241.
- Fadem SZ. The Hemodialysis Procedure. In: *Complications in Dialysis: A Clinical Guide.* Springer; 2023. pp. 1-25.
- Groth T, Wiese J, Schuchardt M, Hoyer PF, *et al* Wearable and implantable artificial kidney devices for end-stage kidney disease treatment: current status and review. *Artif Organs.* 2023;47(4):649-666.
- Sanchez K. Kidney Disorders and Therapeutic Management. *Priorities in Critical Care Nursing-E-Book.* 2022. p. 373.
- Kalariya Y, Gupta A, Patel M. Integrative medicine approaches: bridging the gap between conventional and renal complementary therapies. *Cureus.* 2023;15(9).
- Ang YTI, Gan SWS, Liow CH, *et al* Patients' perspectives of home and self-assist haemodialysis and factors influencing dialysis choices in Singapore. *Ren Replace Ther.* 2022;8(1):42.
- Huang M, Feng Y, Liu J, *et al* Pathways to improve nephrologist comfort in managing patients on in-center or home self-care dialysis. *Hemodial Int.* 2023;27(3):204-211.
- Himmelfarb J, Vanholder R, Mehrotra R, Tonelli M. The current and future landscape of dialysis. *Nat Rev Nephrol.* 2020;16(10):573-585.
- Bodington R, Kassianides X, Bhandari S. Point-of-care testing technologies for the home in chronic kidney disease: a narrative review. *Clin Kidney J.* 2021;14(11):2316-2331.
- Haroon S, Davenport A. Haemodialysis at home: review of current dialysis machines. *Expert Rev Med Devices.* 2018;15(5):337-347.
- Hojs N, Fissell WH, Roy S. Ambulatory hemodialysis-technology landscape and potential for patient-centered treatment. *Clin J Am Soc Nephrol.* 2020;15(1):152-159.
- Adebisi B, Aigbedion E, Ayorinde OB, Onukwulu EC. A Conceptual Model for Predictive Asset Integrity Management Using Data Analytics to Enhance Maintenance and Reliability in Oil & Gas Operations. 2021.
- Paul PO, Abbey ABN, Onukwulu EC, Agho MO, Louis N. Integrating procurement strategies for infectious disease control: Best practices from global programs. *Prevention.* 2021;7:9.
- Adebisi B, Aigbedion E, Ayorinde OB, Onukwulu EC. *International Journal of Social Science Exceptional Research.* 2022.
- Fredson G, Adebisi B, Ayorinde OB, Onukwulu EC, Adediwin O, Ihechere AO. Strategic Risk Management in High-Value Contracting for the Energy Sector: Industry Best Practices and Approaches for Long-Term Success. 2023.
- Majebi NL, Drakeford OM, Adelodun MO, Chinyere E. Leveraging digital health tools to improve early detection and management of developmental disorders in children. *World J Adv Sci Technol.* 2023;4(1):25-32.
- Fiemotongha JE, Igwe AN, Ewim CP-M, Onukwulu EC. Innovative trading strategies for optimizing profitability and reducing risk in global oil and gas markets. *J Adv Multidiscip Res.* 2023;2(1):48-65.
- Fiemotongha JE, Igwe AN, Ewim CP-M, Onukwulu EC. *International Journal of Management and Organizational Research.* 2023.
- Daramola OM, Apeh CE, Basiru JO, Onukwulu EC, Paul PO. Optimizing Reverse Logistics for Circular Economy: Strategies for Efficient Material Recovery and Resource Circularity. 2023.
- Farooq A, Abbey ABN, Onukwulu EC. Optimizing Grocery Quality and Supply Chain Efficiency Using AI-Driven Predictive Logistics. 2023.
- Basiru JO, Ejiofor CL, Onukwulu EC, Attah RU. Sustainable procurement in multinational corporations: A conceptual framework for aligning business and environmental goals. *Int J Multidiscip Res Growth Eval.* 2023;4(1):774-87.
- Basiru JO, Ejiofor L, Onukwulu C, Attah RU. Corporate health and safety protocols: A conceptual model for ensuring sustainability in global operations. *Iconic Res Eng J.* 2023;6(8):324-343.
- Basiru JO, Ejiofor CL, Onukwulu EC, Attah RU. Financial management strategies in emerging markets: a review of theoretical models and practical applications. *Magna Sci Adv Res Rev.* 2023;7(2):123-40.
- Basiru JO, Ejiofor CL, Onukwulu EC, Attah RU. The impact of contract negotiations on supplier relationships: A review of key theories and frameworks for organizational efficiency. *Int J Multidiscip Res Growth Eval.* 2023;4(1):788-802.
- Basiru JO, Ejiofor CL, Onukwulu EC, Attah R. Enhancing financial reporting systems: A conceptual

- framework for integrating data analytics in business decision-making. *IRE J.* 2023;7(4):587-606.
33. Basiru JO, Ejiofor CL, Onukwulu EC, Attah RU. Optimizing administrative operations: A conceptual framework for strategic resource management in corporate settings. *Int J Multidiscip Res Growth Eval.* 2023;4(1):760-73.
 34. Adekola AD, Alli OI, Mbata AO, Ogbeta CP. Integrating Multisectoral Strategies for Tobacco Control: Evidence-Based Approaches and Public Health Outcomes. 2023.
 35. Fredson G, Adebisi B, Ayorinde OB, Onukwulu EC, Adediwin O, Ihechere AO. *International Journal of Social Science Exceptional Research.* 2022.
 36. Onukwulu EC, Fiemotongha JE, Igwe AN, Ewim CP-M. *International Journal of Management and Organizational Research.* 2022.
 37. Basiru JO, Ejiofor CL, Onukwulu EC, Attah RU. Streamlining procurement processes in engineering and construction companies: a comparative analysis of best practices. *Magna Sci Adv Res Rev.* 2022;6(1):118-35.